



Certificate of Compliance

Certificate: 2507678 **Master Contract:** 163467
Project: 2507678 **Date Issued:** April 12, 2012
Issued to: Spellman High Voltage Electronics
Corporation
475 Wireless Blvd
Hauppauge, NY 11788
USA
Attention: Godrej Mehta

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



David Jarvis

Issued by: David Jarvis

PRODUCTS

CLASS 8780 81 - MEDICAL ELECTRICAL EQUIPMENT/SYSTEMS - Certified to US Standards

CLASS 8780 01 - MEDICAL ELECTRICAL EQUIPMENT/SYSTEMS

High Voltage Generator, Model CRX500LX3338, permanently installed rated: 480 Vac, 50/60 Hz, 52 A pulsed, 120Vac, 50/60Hz, 4A. consists of the following:

Cathode model 405791-012 output: -75kV Max, 500mA Max, 30kW max

Filament: 15V @ 7A max.

Grid (G1, G2) -3000Vmax, 4kHz max, 300pfd load max

Anode model 405791-011: +75kV Max, 500mA Max, 30kW Max.

HSS: 190V @ 11A 5 sec max.

280V @ 6.2A, 40 Sec max

210V @ 4A continuous;